

我该如何校正记录器的湿度功能呢？

答：湿度传感器校准仅在在出厂前便已完成。而由于记录器的湿度校正需要特别的工具，所以并非能由使用者端自行校正。

可以联系您购买此数据记录器的商店以获取更多服务。

Q: How to calibrate the humidity of my AZ datalogger?

A: The humidity sensor calibration is done in factory only. Users can not recalibrate it because the special tooling is required. Contact the store where you bought this data logger for further services.