



Worldwide Testing Services (Taiwan) Co., Ltd.

ACCREDITED TEST HOUSE

CERTIFICATION OF TESTING

Under EU EMC - DIRECTIVE 2004/108/EC -

This certifies that the following designated product

Dissolved Oxygen Meter

Model No. : 8403

Multi-listing Model No. : 8401, 8402

Brand Name : AZ

.....
(Product identification)

Has been tested in accordance to essential protection requirements of Council Directive 2004/108/EC and found the test results indeed meet the limitation of the relevant test standard(s) listed below:

EN 61326-1: 2006

**(CISPR11, IEC/EN 61000-3-2 (2006), IEC/EN 61000-3-3 (2008),
IEC/EN61000-4-2 (1995+A1:1998+A2: 2001)/-3 (2006+A1:2008)/-4(2004)
/-5(2006)/-6 (2007)/-11 (2004))**

.....
(Identification of regulations / standards)

This certificate is issued for

AZ Instrument Corporation

No.3-2, Chien-Kuo road, TEPZ Tantz, Taichung 427 Taiwan R.O.C.

.....
(Name / Address)

SPECIAL STATEMENT:

1. THE OPERATION CONDITIONS OF THE TEST SAMPLE DURING THE ALL TEST ITEMS WERE UNDER THE DEMAND OF THE APPLICANT. ANY DEVIATION FROM THE REQUIRED CONDITION OF TEST STANDARD IS THE RESPONSIBILITY OF THE APPLICANT.
2. THE CERTIFICATION IS VALID ONLY IN CONNECTION TO THE TEST REPORT NUMBER W6R20905-9761-E-11 AND TO THE SAMPLE HAS BEEN TESTED BY WORLDWIDE TESTING SERVICES (TAIWAN) CO., LTD.

June 29, 2009

.....
(Date)

Chang Tse-ming

.....
Chang Tse-Ming, Laboratory Director

