

**DECLARATION OF IDENTICAL TO TYPE**

Hereby we

**AZ INSTRUMENT CORPORATION**

(Name of manufacturer)

No. 20 23<sup>rd</sup> road Taichung Industrial park , Taichung City, Taiwan R.O.C.

Tel:886-4-23591508

Fax:886-4-23593991

(Details of address, telephone number and fax number )

**declare that the equipment**

New name : **Hygro-thermometer**

New model number of product : **8723**

**is structurally identical to the basic model**

**Thermohygrometer Model No. : AZ8721**

**G5R201120247-E-16** (reference no. of test report / BZT -type approval no.)

Regulations applied : EN 61326-1: 1997, EN 55022, EN 61000-4-2/-3

Test specifications : Emissions: EN 55022:1998

ESD: EN 61000-4-2 : 1996

RF-Field: EN61000-4-3:1999



2006/10/25

Authorized Signature



