



ELECTRONIC TECHNOLOGY SYSTEMS
DR. GENZ GMBH

ACCREDITED TEST HOUSE - LVD, GS-MARK, TOY SAFETY -
NOTIFIED BODY / GS-CERTIFICATION BODY

EC DECLARATION OF CONFORMITY

- LOW VOLTAGE DIRECTIVE -
Page 1 of 2

This certifies that the following designated product

**INFRARED THERMOMETER
MODEL NO. : AZ8866**

.....
(Product identification)

complies with the requirements of the European Community Directive 73/23/EEC and 93/68/EEC.
This declaration applies to all specimens manufactured in accordance with the attached
manufacturing drawings which form part of this declaration.

Assessment of compliance of the product with the requirements relating to the Low Voltage
Directive (LVD) was based on the following standards:

EN 60825-1

.....
(Identification of regulations / standards)

This declaration is the responsibility of the manufacturer / importer

**AZ INSTRUMENT CORP.
NO. 20 23RD ROAD, TAICHUNG INDUSTRIAL PARK
TAICHUNG CITY, TAIWAN, R.O.C**

.....
(Name / Address)



THE CERTIFICATION IS ONLY VALID WITH THE TESTREPORT NR. G5M199080045-L-01
THE CONFIRMATION IS ONLY VALID IF THE NON-CONFIRMATION ITEMS ARE ELIMINATED

MANUFACTURER / IMPORTER

TEST LABORATORY

This is the result of test, that was carried out from the
submitted type-samples of a product in conformity with
the specification of the respective standards.
The certificate holder has the right to fix the CE-mark for
LVD on the product complying with the inspection sample.

AUG. 30 1999

.....
(Date)

Cheng Shan Heng
AZ Instrument Corp.
Taichung Taiwan
.....
(Surname, forename)
.....
(Company, address)
AZ Instrument Corp., No. 20, 23rd Road, Taichung Industrial Park, Taichung City, Taiwan, R.O.C.



August 30, 1999

.....
(Date)

Dr. Genz

.....
Dr. Genz

ELECTRONIC TECHNOLOGY SYSTEMS DR. GENZ GMBH

DR. GENZ GMBH, D-15526 REICHENWALDE B. BERLIN, GERMANY, PHONE +49-33631-88800 FAX +49-33631-88866



ELECTRONIC TECHNOLOGY SYSTEMS
DR. GENZ GMBH

ACCREDITED TEST HOUSE - LVD, GS-MARK, TOY SAFETY -
NOTIFIED BODY / GS-CERTIFICATION BODY

EC DECLARATION OF CONFORMITY

- LOW VOLTAGE DIRECTIVE -
Page 2 of 2

Test report Nr. G5M199080045-L-01
Date: August 30, 1999

The table below is a summary of the number of non-compliances of each grade

Grade:	Classification	Number
Remark	A	1
Severe non-compliance	B	1
Very severe non-compliance	C	-

Non-compliance	Grade
Marking incomplete (type reference, mark of origin)	A
Marking incomplete (safety marking)	B
-over-	



ELECTRONIC TECHNOLOGY SYSTEMS DR. GENZ GMBH

STORKOWER STRASSE 38C, D-15526 REICHENWALDE B. BERLIN, GERMANY, PHONE + 49-33631-88800 FAX + 49-33631-88866